

SYRINGE PUMP BATTERY

TE 311-312-332

TERUMO INDIA PRIVATE LIMITED

Tower B, 16TH FLOOR, UNITECH CYBER PARK, UNITECH WORLD,
JHASA, GURGAON GURUGRAM, HARYANA

Phone: +911244718700, +911244718718 Email:
Neha_ahuja@terumo.co.jp
<http://www.hospYTEK.com/terumo-india-private-limited>

OVERVIEW:

DESCRIPTION:

Battery Type: NI-MH Battery Volt: 9.6V Battery Capacity: 2000mAh Battery
Color: Blue Battery Weight: 188g Battery Length: 58*30*52mm Battery Warranty: 12
months This Medical Instrument Battery can replace the following part
numbers: BN-600AAK This Medical Instrument battery is also compatible with the
following model: TE-331, TE-311, TE-312, TE-332 Overcurrent overcharge and over
discharge circuit protection, temperature-controlled all-copper imports, import
terminals



TECHNICAL SPECIFICATIONS:

Chemistry:	NiCD	Voltage:	9.6V	Nominal
Capacity:	1.1Ah	Device Manufacturer:	Explore Synergy Inc	Application:
	Terumo TE-311-332			

READ MORE